

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 08 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130173
Name <u>J Hemburg</u>		ID#	5828
Date		9/24/09	
A	Agency	AST	Phone # 822-3263
Instrument Location <u>Glennallen</u> ✓			
Alco S/N		X124625	Target Value 0.083% ✓
High Pressure		1200	
B	Alco Test Values	0.079 ✓	0.079 ✓
		1 st Alco	2 nd Alco
Signature		9/24/09	
(OVER)			

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 28 day of Oct, 2009.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

13173
130173

98-09
CIV

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

Breath Test Record

Serial Number 130173 Model DataMaster
Location GKN0 AST/Glennallen

9/25/09

Subject Information

Last Name Voc O.L. # VOC
First Name Voc

Operator Information

Operator Name Heinbaugh Case VOC
Operator # 5828 Department GKN0

Test Results

Alco Target 0.083 Alco Serial X124625
Nominal 0.078 Barometer 28.40 Test Type V
Dry Gas 1 0.079 Time: 11:20:34 Test Date 09/04/2009
Subject Alcohol 0.000 Time: 11:21:47 Test Time 11:18:52
Dry Gas 2 0.079 Time: 11:23:02 Report Y

DIAGNOSTIC CHECK

Diagnostic Test Record

Serial Number 130173 Model DataMaster
Location GKN0 AST/Glennallen

Diagnostic Test Results

Test Date 09/04/2009 Computer Okay
Test Time 11:24:00 Program Okay
Software Date 02/20/2001
Sample Chamber (C) 49
Breath Tube (C) 42
Barometer (in) 28.40
Flow Detector Okay
High Speed Pump Okay
Detector Okay
Filter Okay
Quartz Okay
Calibration Okay